

## Datasheet

# Pancoll

## Separating solutions

Product	Description	Catalogue-No.	Size
Pancoll human	Separating Solution, Density: 1.077 g/ml	P04-60100	100 ml
		P04-60500	500 ml
Pancoll animal	Separating Solution, Density: 1.077 g/ml	P04-63100	100 ml
		P04-63500	500 ml
Pancoll mouse	Separating Solution, Density: 1.086 g/ml	P04-64100	100 ml
		P04-64500	500 ml
Pancoll rat	Separating Solution, Density: 1.091 g/ml	P04-65100	100 ml
		P04-65500	500 ml
Pancoll Platelets	Separating Solution, Density: 1.063 g/ml	P04-67100	100 ml
		P04-67500	500 ml
Pancoll monocytes	Separating Solution, Density: 1.068 g/ml	P04-68100	100 ml
		P04-68500	500 ml
Pancoll human	Separating Solution, Density: 1.100 g/ml	P04-69610	100 ml
		P04-69600	500 ml

### Product description

Pancoll separating solutions from PAN-Biotech contain a polysaccharide with a molecular weight of 400,000 daltons; this hydrophilic polymer allows for production of aqueous solutions for cell separation with a density of up to 1.2 g/ml. PAN-Biotech offers a variety of ready-to-use products with a density of 1.063 g/ml up to 1.091 g/ml for a very wide range of cell separation applications.

### Storage conditions

Storage: +2 to +20°C, in the dark  
 Stability: 3 years from date of production  
 Size: 100 ml, 500 ml, other sizes on request

### Method of Separation

For lymphocyte separation blood is used which has been defibrinated or treated with anticoagulants (Heparin, EDTA, Citrate), and which is diluted one to three-fold, depending on the haematocrit level of the blood sample, with the required volume of a physiological saline solution. Then the Pancoll solution is carefully covered with a layer of diluted blood in a centrifuge vial, without mixing the phases. After a short centrifugation step (e.g. 800 g for 20 minutes) at room temperature the lymphocytes, together with monocytes and platelets, can be harvested from the white blood cells layer between the plasma sample layer and the Pancoll. The separated cells are then washed twice in physiological saline solution to purify the lymphocytes by removing platelets.

During centrifugation the cells of the blood sample migrate to the Pancoll layer where they get into contact with the polysaccharide contained in Pancoll. The red blood cells are aggregated by this substance at room temperature immediately. Aggregation causes an increase of the sedimentation rate of the red blood cells which aggregate together with the granulocytes as a sediment at the bottom of the centrifuge vial. Lymphocytes, monocytes and platelets are not so dense and cannot enter and pass through the Pancoll layer. These cells are concentrated as white blood cell layer above the Pancoll layer and therefore can be harvested easily by careful pipetting.

In subsequent centrifugation steps the lymphocytes are washed to remove remaining platelets, serum and Pancoll. As a result of this process a highly purified suspension of viable lymphocytes and monocytes (PBMC) is obtained.

### Sample preparation

Blood samples should be processed as soon as possible after they have been obtained in order to achieve optimum results and cell viability.

Storing blood samples at room temperature for more than 12 hours will cause a reduced yield of lymphocytes, a change in the surface markers and an impaired response to mitogen stimulation.

### Protocol for Isolating Lymphocytes

- Transfer Pancoll at ambient temperature under sterile conditions into a suitable sterile centrifuge tube. Alternatively 50 ml tubes, prefilled with 15 ml Pancoll (P04-60125) or 10 ml tubes prefilled with 3 ml Pancoll (P04-60225) can be used. The 50 ml tubes are used with 15-30 ml blood samples and 10 ml tubes are used with 3-8 ml blood samples.
- Carefully cover the separating solution with a layer of the blood sample which is diluted one to three-fold with physiologic saline solution.
- **Important: Do not mix the blood sample with Pancoll!**
- Centrifuge at 800 g at 20°C for 20 minutes. **SWITCH BRAKE OFF!**
- After centrifugation, carefully remove the upper phase (containing plasma and platelets) using a pipette, without mixing the interphase with the lymphocytes.
- Using a new pipette, transfer the lymphocyte band above the Pancoll-layer to a new centrifuge vial. It is important to remove all the lymphocyte band of the interphase with as little volume as possible. If too much Pancoll (lower phase) is picked up, a contamination with granulocytes may occur. If too much supernatant (upper phase) is picked up an increased contamination with platelets will occur.
- Add at least 3 volumes of a physiological saline solution to the lymphocytes.
- Suspend the lymphocytes carefully using a pipette.
- Centrifuge at 300 g at 20°C for 10 minutes.
- Discard the supernatant.
- Repeat the washing step two times.

### Typical results with Pancoll

Lymphocytes	60 ± 20 %	yield of Lymphocytes from original blood samples
	95 ± 5 %	of the Lymphocyte fraction are mononuclear Leukocytes
	> 90 %	live cells (trypan blue-exclusion)
Other cells	3 ± 2 %	Granulocytes
	5 ± 2 %	Erythrocytes
	< 0.5 %	total number of platelets of the original blood sample

Result:	Possible cause:	Comment:
Contamination of the Lymphocyte fraction with Erythrocytes and Granulocytes	<ul style="list-style-type: none"> <li>• Temperature too low</li> <li>• Centrifugation speed too low and/or time too short</li> </ul>	The density of Pancoll is higher at lower temperatures, the Erythrocytes aggregate less and they cannot penetrate (also the Granulocytes) the Pancoll properly <b>Increase Pancoll temperature to 20°C</b>
Low yield and viability of Lymphocytes	<ul style="list-style-type: none"> <li>• Temperature too high</li> </ul>	Adequate times and G-forces have to be kept to assure a complete sedimentation of non-lymphoid cells. Pancoll has a lower density at higher temperatures and Lymphocytes can penetrate to Pancoll easier
Low yield of Lymphocytes with normal viability	<ul style="list-style-type: none"> <li>• Blood sample not diluted with buffer</li> <li>• Abnormal high haematocrit in blood sample</li> </ul>	At very high cell densities Lymphocytes can be included in aggregates of Erythrocytes <b>Dilute the blood sample</b>
Low yield of Lymphocytes with contamination of Granulocytes	<ul style="list-style-type: none"> <li>• Vibrations of the centrifuge rotor can disturb the gradient</li> </ul>	Vibrations can result in a broadening of the Lymphocyte band and to a stirring with the cells below <b>Balance the rotor and switch-off the brake of the centrifuge</b>
Low yield of Lymphocytes with contamination of other cell types	<ul style="list-style-type: none"> <li>• Sample contains cells with abnormal densities</li> </ul>	Can happen with pathologic blood samples or with samples of non-peripheral blood

### Technical support

For technical support, questions or remarks please contact your local PAN-Biotech partner or the technical department of PAN-Biotech via email ([info@pan-biotech.com](mailto:info@pan-biotech.com)) or phone +49-8543-601630.

---

FOR RESEARCH USE ONLY! Not approved for human or animal diagnostic or therapeutic procedures.